

# wile<sup>plus</sup> Rice

Moisture meter



EN Operators manual

# Box Contents

Thank you for choosing Wile Rice moisture meter. The meter ensures quick and easy measurement of moisture, temperature and hectoliter weight of the sample. Automatic temperature and hectoliter weight compensation guarantees accurate results. The meter is equipped with a patented integrated Rice leveling system. Excess Rice goes into the compartment around test cell keeping working area neat during whole measuring process. Read this manual carefully to learn how to operate this unit correctly.

- 1 METER
- 2 CARRYING CASE
- 3 USB CABLE
- 4 BATTERIES
- 5 MANUAL



Important: The scales of Wile moisture meters are developed to measure average quality grain. Exceptional growth conditions and new varieties of Rice can considerably affect Rice properties. Therefore, in the beginning of new harvest season we recommend to check the readings of your meter against an oven dried sample.

## FUNCTIONS

### SYMBOL



<4.0

>28.0



### Definition

Replace batteries



Moisture is below the lower limit for this grain


Moisture is above the upper limit for this grain

Error code

## MEASURING PROCESS

Wile guides user step by step directly on the screen during the whole measuring process.

1. Place hopper on tester. Press and release ON/OFF  button to turn on tester. After power up the display shows the results of the last Rice tested.
2. Select Rice using the UP/DOWN buttons. Once the desired Rice is selected then press the ENTER button. Place tester on a stable and mostly level surface. Press ENTER to initiate test. 



3. Ensure hopper slider is pushed in to keep Rice from dropping. Fill hopper to the rim with sample.
4. Pull slider and the rice will evenly flow into the test cell.
5. Remove hopper when prompted. Level the rice by pushing the handle in steady motion towards the back of the unit and back removing excess grain.
6. When the handle is back to starting position, the meter will automatically start measuring. Results will appear on the screen within seconds.




## FUNCTIONS AND SETTINGS

Settings and functions can be browsed and modified at the settings menu



### RICE OFFSET (GRAIN OFFSET)

Each Ricecan be individually adjusted by 10.0% in increments of 0.1% to match the results of an elevator tester




### LANGUAGE

Select preferred language




## UNIT

Select preferred weight (kg/hl or lb/Bu) and temperature units (Celsius or Fahrenheit). When the temperature units are changed from Fahrenheit to Celsius the test weight units are also changed




## AUTO OFF

The meter will turn off automatically after a selected time (30s, 1min, 5min, 10min, 20min)




## AVERAGE CALCULATION

The meter automatically calculates the average of the previous measurements. In settings menu you can adjust the number of results (3, 6 or 9 measurements) for the average calculation. The average can also be reset.




## BACKLIGHT ON/OFF

To extend battery life you can turn display backlight off.




## LCD CONTRAST

Adjust LCD display contrast.



## BATTERY


Battery level can be checked. If the battery level is under 22% the meter will indicate it on the startup screen automatically, but measuring is still possible. Low battery voltage does not affect the measurement results.



## WEIGHT TEST


Scale function can be checked at weight test mode.

Use known weight sample up to 500 grams of weight. Follow instruction on the screen of your meter.




## INFO

Info section shows software version details.



**CLEANING THE TESTER — TEST CELL**

Use a soft, clean cloth to wipe out the cell, using extreme care not to damage the Rice temperature sensor at the bottom of the test cell.

**MAINTENANCE**

Use a soft, clean cloth to clean device. Do not use strong detergents and do not allow liquids inside the meter. Keep the meter in a dry place, preferably at room temperature. If the meter is not used for a long time, remove the batteries.
















**BATTERY REPLACEMENT**

The meter uses four AA (LR6) batteries. Replace all batteries at the same time.



# Error Codes

Wile moisture meter is equipped with error messages to guide the user to get the most reliable results, as well as to ensure that the meter is in good condition.

Error code	Cause	Solution
 #	Critical Error	The meter must be submitted to an authorized service center
 #7	Weight not detected	Repeat test / Do weight test
 #8	Taring result is too high	Do weight test and check the scale
 #9	Weight measurement uncertain / meter shook during the measurement	Keep the meter stable, measurement uncertainty has increased
 #10	Moisture measurement result stabilize slowly	Happens to very wet samples
 #11	Weight measurement too uncertain / meter shook during the measurement	Keep the meter stable, measurement uncertainty has increased
 #12	Temperature of the sample lower than 2 C / Measurement uncertainty is high	The tester is most accurate when Rice and tester temperature is between 16C and 32C
 #13	Temperature of the sample above 50 C / Measurement uncertainty is high	The tester is most accurate when Rice and tester temperature is between 16C and 32C
 #14	Large temp. diff. between the Rice and test cell / Measurement uncertainty has increased	For better accuracy, Rice and meter test cell should be at similar temperatures
 #15	Not on a sufficiently level surface, meter warns with sound alarm	Keep meter stable during measurement
 #16	Weight result is over 330 g	Check that leveling handle did not stay above the test cell/ Do weight test
 #17	Weigh result is under 50 g	Check that leveling handle did not stay above the test cell/ Do weight test
 #18	Hectoliter weight is 20% higher than average	Measurement uncertainty has increased, Do weight test
 #19	Hectoliter weight is 20% lower than average	Measurement uncertainty has increased, Do weight test
 #20	Test cell is not cleared before the measurement	Clear test cell



This product has a warranty valid for one (1) year from the date of purchase. The warranty covers the materials and workmanship.

To claim the warranty, the customer should return the defected product to the Manufacturer, reseller or the nearest Wile Service Partner at the customer's own expense. The warranty claim must be accompanied with the description of the fault, copy of the sales receipt and customer's contact information.

The manufacturer / Service Partner will repair or replace the defected product and return it as soon as possible.

The warranty does not cover any damages that are caused by incorrect or careless use of the product, installation that does not correspond to the provided instructions and other damages which may arise due to causes beyond the control of the manufacturer.

The liability of the manufacturer is limited to the pRice of the product at a maximum.

Manufacturer does not accept any responsibility for any direct, indirect or consequential damages that are caused by the use of the product or the fact that the product could not be used.

Wearing parts are not covered by the warranty e.g. Battery, Case.







Jusslansuora 8, 04360  
TUUSULA, FINLAND  
Tel. +358 9 7744 970  
[info@farmcomp.fi](mailto:info@farmcomp.fi)  
[www.wile.fi](http://www.wile.fi)

**ERC**